Amplivox Otowave 102, 202, 302 – Items required for calibration (items manufactured by *G.R.A.S., Digitron & EXTECH* are shown as examples)



Sound Level Meter
(G.R.A.S. HW1001)



Microphone calibrator
(G.R.A.S. 42AB)



Microphone
(G.R.A.S. 40AG)



2cc Coupler
(G.R.A.S. RA0038)



Microphone preamplifier/cable
(G.R.A.S. 26AM)



Pressure Meter +/-2 PSI
(Digitron PM-80 or EXTECH HD700)



Calibration Cavities
(Amplivox T027, T028, T029 + T042 for 202/302)



Pressure Adapter
(Amplivox CAL.OTO)



Ear tip 11mm Umbrella
(Amplivox 8002013)



Service Manual
(Amplivox MANSERV)